

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010860.D
 Acq On : 07 Oct 2022 10:35
 Operator : KP/MD
 Sample : VY1007SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1007SBL01

Quant Time: Oct 08 04:45:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	384608	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	339527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	148182	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	130985	58.356	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.720%
35) Dibromofluoromethane	7.716	113	123238	48.271	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.540%
50) Toluene-d8	10.179	98	459975	52.682	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.360%
62) 4-Bromofluorobenzene	12.483	95	139541	40.185	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
Data File : VY010860.D
Acq On : 07 Oct 2022 10:35
Operator : KP/MD
Sample : VY1007SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1007SBL01

Quant Time: Oct 08 04:45:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

