

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010866.D
 Acq On : 07 Oct 2022 13:00
 Operator : KP/MD
 Sample : N5004-20RE
 Misc : 6.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D22012-18RE

Quant Time: Oct 08 04:46:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	265241	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	459938	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	339063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	101377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	389	0.145	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.280%#
35) Dibromofluoromethane	7.722	113	268m	0.088	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.180%#
50) Toluene-d8	10.179	98	1816	0.174	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.340%#
62) 4-Bromofluorobenzene	12.483	95	1290	-3.630	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	-7.260%#
Target Compounds						
36) 1,1-Dichloropropene	7.783	75	21072	4.781	ug/l #	50
81) trans-1,4-Dichloro-2-b...	12.483	75	755	1.536	ug/l #	7

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
Data File : VY010866.D
Acq On : 07 Oct 2022 13:00
Operator : KP/MD
Sample : N5004-20RE
Misc : 6.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D22012-18RE

Quant Time: Oct 08 04:46:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

