

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010868.D
 Acq On : 07 Oct 2022 13:49
 Operator : KP/MD
 Sample : N5004-14RE
 Misc : 6.84g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 D22013-15.5RE

Quant Time: Oct 08 04:46:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	175698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	311406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	254948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	82258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	48359	27.154	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	54.300%
35) Dibromofluoromethane	7.716	113	40838	19.756	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	39.520%#
50) Toluene-d8	10.179	98	97073	13.731	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	27.460%#
62) 4-Bromofluorobenzene	12.483	95	33532	9.134	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	18.260%#
Target Compounds						
16) Acetone	3.942	43	9323	20.612	ug/l	96
19) Methyl tert-butyl Ether	5.216	73	17926	4.122	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
Data File : VY010868.D
Acq On : 07 Oct 2022 13:49
Operator : KP/MD
Sample : N5004-14RE
Misc : 6.84g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
D22013-15.5RE

Quant Time: Oct 08 04:46:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

