

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010877.D
 Acq On : 07 Oct 2022 17:13
 Operator : KP/MD
 Sample : N5046-07
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-7

Quant Time: Oct 08 04:48:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	144235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	252800	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	229343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	92513	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	93595	64.018	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.040%
35) Dibromofluoromethane	7.716	113	87204	51.966	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.940%
50) Toluene-d8	10.179	98	308036	53.675	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.340%
62) 4-Bromofluorobenzene	12.477	95	90349	39.526	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
Data File : VY010877.D
Acq On : 07 Oct 2022 17:13
Operator : KP/MD
Sample : N5046-07
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-7

Quant Time: Oct 08 04:48:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

