

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010880.D
 Acq On : 07 Oct 2022 18:21
 Operator : KP/MD
 Sample : N5046-11
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-11

Quant Time: Oct 08 04:48:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	214409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	377290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	334886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	143001	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	124450	57.263	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.520%
35) Dibromofluoromethane	7.710	113	118530	47.327	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.660%
50) Toluene-d8	10.179	98	437303	51.057	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.120%
62) 4-Bromofluorobenzene	12.477	95	132434	38.749	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
Data File : VY010880.D
Acq On : 07 Oct 2022 18:21
Operator : KP/MD
Sample : N5046-11
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-11

Quant Time: Oct 08 04:48:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration

