

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010874.D
 Acq On : 07 Oct 2022 16:05
 Operator : KP/MD
 Sample : N5046-02
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 R-2

Quant Time: Oct 08 04:47:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 04 07:26:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	255698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	454972	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	404050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	169282	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	151307	58.379	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.760%
35) Dibromofluoromethane	7.716	113	145236	48.089	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.180%
50) Toluene-d8	10.179	98	535558	51.852	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.700%
62) 4-Bromofluorobenzene	12.477	95	161994	39.363	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.720%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
Data File : VY010874.D
Acq On : 07 Oct 2022 16:05
Operator : KP/MD
Sample : N5046-02
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
R-2

Quant Time: Oct 08 04:47:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 04 07:26:22 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05




Instrument :
MSVOA_Y
ClientSampleId :
R-2


Tgt Ion:168 Resp: 255698
Ion Ratio Lower Upper
168 100
99 56.0 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 58.379 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.006 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05

Tgt Ion: 65 Resp: 151307
Ion Ratio Lower Upper
65 100
67 54.5 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05




Instrument :
MSVOA_Y
ClientSampleId :
R-2


Tgt Ion:114 Resp: 454972
Ion Ratio Lower Upper
114 100
63 20.5 0.0 43.2
88 15.6 0.0 31.8



#35
Dibromofluoromethane
Concen: 48.089 ug/l
RT: 7.716 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05

Tgt Ion:113 Resp: 145236
Ion Ratio Lower Upper
113 100
111 102.9 82.6 124.0
192 19.3 17.6 26.4









#50
Toluene-d8
Concen: 51.852 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05




Instrument :
MSVOA_Y
ClientSampleId :
R-2


Tgt Ion: 98 Resp: 535558
Ion Ratio Lower Upper
98 100
100 64.8 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 39.363 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05

Tgt Ion: 95 Resp: 161994
Ion Ratio Lower Upper
95 100
174 83.6 0.0 177.0
176 79.3 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05

Instrument :
MSVOA_Y
ClientSampleId :
R-2

Tgt Ion:117 Resp: 404050
Ion Ratio Lower Upper
117 100
82 54.4 43.0 64.4
119 32.1 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. 0.000 min
Lab File: VY010874.D
Acq: 07 Oct 2022 16:05

Tgt Ion:152 Resp: 169282
Ion Ratio Lower Upper
152 100
115 59.2 28.2 84.7
150 158.7 0.0 347.6

