



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY100722\
 Data File : VY010858.D
 Acq On : 07 Oct 2022 08:17
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100322S.M
 Title : SW846 8260
 Last Update : Tue Oct 04 07:26:22 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	20331	PASS
75	95	30	60	50.8	55301	PASS
95	95	100	100	100.0	108845	PASS
96	95	5	9	6.4	6990	PASS
173	174	0.00	2	1.1	989	PASS
174	95	50	100	85.0	92525	PASS
175	174	5	9	7.4	6833	PASS
176	174	95	101	96.2	89008	PASS
177	176	5	9	6.4	5728	PASS