

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101624\
 Data File : VY019914.D
 Acq On : 16 Oct 2024 14:44
 Operator : SY/MD
 Sample : P4410-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-101524

Quant Time: Oct 17 01:52:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	224722	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	453828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	407871	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	144539	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	161735	58.452	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.900%
35) Dibromofluoromethane	7.634	113	149087	49.783	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.560%
50) Toluene-d8	10.109	98	556374	50.308	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.620%
62) 4-Bromofluorobenzene	12.407	95	171666	42.959	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	85.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101624\
Data File : VY019914.D
Acq On : 16 Oct 2024 14:44
Operator : SY/MD
Sample : P4410-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-101524

Quant Time: Oct 17 01:52:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

