

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011086.D
 Acq On : 25 Oct 2022 10:45
 Operator : KP/MD
 Sample : VY1025SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1025SBL01

Quant Time: Oct 26 08:36:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	221677	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	374212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	330168	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	151670	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	101651	49.781	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.560%
35) Dibromofluoromethane	7.716	113	106308	51.007	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.020%
50) Toluene-d8	10.179	98	342904	43.886	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.780%
62) 4-Bromofluorobenzene	12.477	95	120907	43.270	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011086.D
Acq On : 25 Oct 2022 10:45
Operator : KP/MD
Sample : VY1025SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1025SBL01

Quant Time: Oct 26 08:36:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

