

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011093.D
 Acq On : 25 Oct 2022 14:56
 Operator : KP/MD
 Sample : N5257-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HOLLAND-RD-COMP-2

Quant Time: Oct 26 08:38:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	84689	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	142026	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	122570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	46457	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	55456	71.087	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 142.180%			
35) Dibromofluoromethane	7.710	113	54392	69.546	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 139.100%			
50) Toluene-d8	10.179	98	157116	52.981	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 105.960%			
62) 4-Bromofluorobenzene	12.477	95	47351	44.754	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 89.500%			
Target Compounds						
16) Acetone	3.948	43	1893	8.356	ug/l	95
20) Methylene Chloride	4.710	84	12945	7.777	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011093.D
Acq On : 25 Oct 2022 14:56
Operator : KP/MD
Sample : N5257-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HOLLAND-RD-COMP-2

Quant Time: Oct 26 08:38:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

