

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011095.D
 Acq On : 25 Oct 2022 15:42
 Operator : KP/MD
 Sample : N5257-07
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HOLLAND-RD-COMP-4

Quant Time: Oct 26 09:00:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	220749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	379350	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	327796	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	138506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	138388	68.057	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.120%	
35) Dibromofluoromethane	7.710	113	145720	69.763	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 139.520%	
50) Toluene-d8	10.173	98	476752	60.190	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 120.380%	
62) 4-Bromofluorobenzene	12.477	95	153857	55.152	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 110.300%	
Target Compounds						
20) Methylene Chloride	4.710	84	36569	9.035	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011095.D
Acq On : 25 Oct 2022 15:42
Operator : KP/MD
Sample : N5257-07
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HOLLAND-RD-COMP-4

Quant Time: Oct 26 09:00:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

