

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011097.D
 Acq On : 25 Oct 2022 16:27
 Operator : KP/MD
 Sample : N5256-05RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-H1-14.5-15.0RE

Quant Time: Oct 26 09:00:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	337309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	598302	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	535120	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	240410	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	136766	44.017	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.040%
35) Dibromofluoromethane	7.716	113	141419	42.061	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.120%
50) Toluene-d8	10.179	98	470300	37.646	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	75.300%
62) 4-Bromofluorobenzene	12.477	95	165644	36.609	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011097.D
Acq On : 25 Oct 2022 16:27
Operator : KP/MD
Sample : N5256-05RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-H1-14.5-15.0RE

Quant Time: Oct 26 09:00:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

