

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011099.D
 Acq On : 25 Oct 2022 17:12
 Operator : KP/MD
 Sample : N5256-09
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-N2-13.5-14.0

Quant Time: Oct 26 09:00:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	234844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	405346	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	362281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	152948	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	92443	42.733	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.460%
35) Dibromofluoromethane	7.710	113	96898	42.564	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.120%
50) Toluene-d8	10.179	98	316074	37.345	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	74.680%
62) 4-Bromofluorobenzene	12.477	95	106491	34.572	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.140%
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	4239	6.748	ug/l	94
20) Methylene Chloride	4.710	84	28360	4.626	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011099.D
Acq On : 25 Oct 2022 17:12
Operator : KP/MD
Sample : N5256-09
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-N2-13.5-14.0

Quant Time: Oct 26 09:00:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

