

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011100.D
 Acq On : 25 Oct 2022 17:35
 Operator : KP/MD
 Sample : N5268-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-03-102422

Quant Time: Oct 26 09:01:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	16303	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	27534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	24563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	8674	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	8014	53.365	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.720%
35) Dibromofluoromethane	7.710	113	4880	30.975	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	61.960%
50) Toluene-d8	10.179	98	20080	34.927	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	69.860%
62) 4-Bromofluorobenzene	12.477	95	6483	30.645	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	61.280%
Target Compounds						
20) Methylene Chloride	4.704	84	2051	5.121	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011100.D
Acq On : 25 Oct 2022 17:35
Operator : KP/MD
Sample : N5268-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-03-102422

Quant Time: Oct 26 09:01:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

