

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011104.D
 Acq On : 25 Oct 2022 19:05
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 26 09:02:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.332	0.284	14.5	69	0.00
3 P	Chloromethane	0.476	0.383	19.5	66	0.00
4 C	Vinyl Chloride	0.539	0.455	15.6#	67	0.00
5 T	Bromomethane	0.373	0.309	17.2	71	0.00
6 T	Chloroethane	0.360	0.322	10.6	72	0.00
7 T	Trichlorofluoromethane	0.789	0.745	5.6	75	0.00
8 T	Diethyl Ether	0.295	0.287	2.7	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.486	1.8	78	0.00
10 T	Methyl Iodide	0.606	0.546	9.9	67	0.00
11 T	Tert butyl alcohol	0.062	0.057	8.1	90	0.00
12 CM	1,1-Dichloroethene	0.474	0.430	9.3#	72	0.00
13 T	Acrolein	0.074	0.055	25.7#	58	0.00
14 T	Allyl chloride	0.752	0.703	6.5	73	0.00
15 T	Acrylonitrile	0.154	0.169	-9.7	86	0.00
16 T	Acetone	0.134	0.122	9.0	72	0.00
17 T	Carbon Disulfide	1.391	1.057	24.0	62	0.00
18 T	Methyl Acetate	0.428	0.405	5.4	85	0.00
19 T	Methyl tert-butyl Ether	1.325	1.390	-4.9	80	0.00
20 T	Methylene Chloride	0.761	0.605	20.5	82	0.00
21 T	trans-1,2-Dichloroethene	0.535	0.502	6.2	74	0.00
22 T	Diisopropyl ether	1.611	1.719	-6.7	81	0.00
23 T	Vinyl Acetate	0.963	1.052	-9.2	80	0.00
24 P	1,1-Dichloroethane	0.973	0.985	-1.2	79	0.00
25 T	2-Butanone	0.203	0.207	-2.0	81	0.00
26 T	2,2-Dichloropropane	0.864	0.818	5.3	75	0.00
27 T	cis-1,2-Dichloroethene	0.617	0.624	-1.1	78	0.00
28 T	Bromochloromethane	0.303	0.369	-21.8	76	0.00
29 T	Tetrahydrofuran	0.125	0.137	-9.6	85	0.00
30 C	Chloroform	0.980	1.035	-5.6#	81	0.00
31 T	Cyclohexane	0.881	0.745	15.4	70	0.00
32 T	1,1,1-Trichloroethane	0.862	0.875	-1.5	79	0.00
33 S	1,2-Dichloroethane-d4	0.461	0.387	16.1	67	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.304	0.253	16.8	70	0.00
36 T	1,1-Dichloropropene	0.465	0.445	4.3	74	0.00
37 T	Ethyl Acetate	0.258	0.291	-12.8	86	0.00
38 T	Carbon Tetrachloride	0.480	0.486	-1.3	78	0.00
39 T	Methylcyclohexane	0.557	0.508	8.8	70	0.00
40 TM	Benzene	1.366	1.356	0.7	77	0.00
41 T	Methacrylonitrile	0.140	0.168	-20.0	102	0.01
42 TM	1,2-Dichloroethane	0.378	0.396	-4.8	80	0.00
43 T	Isopropyl Acetate	0.460	0.499	-8.5	82	0.00
44 TM	Trichloroethene	0.370	0.364	1.6	77	0.00
45 C	1,2-Dichloropropane	0.347	0.369	-6.3#	81	0.00
46 T	Dibromomethane	0.193	0.208	-7.8	83	0.00
47 T	Bromodichloromethane	0.466	0.510	-9.4	83	0.00
48 T	Methyl methacrylate	0.206	0.225	-9.2	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
 Data File : VY011104.D
 Acq On : 25 Oct 2022 19:05
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 26 09:02:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	87	0.00
50 S	Toluene-d8	1.044	0.850	18.6	63	0.00
51 T	4-Methyl-2-Pentanone	0.257	0.296	-15.2	87	0.00
52 CM	Toluene	0.848	0.872	-2.8#	78	0.00
53 T	t-1,3-Dichloropropene	0.484	0.512	-5.8	80	0.00
54 T	cis-1,3-Dichloropropene	0.554	0.585	-5.6	80	0.00
55 T	1,1,2-Trichloroethane	0.279	0.309	-10.8	85	0.00
56 T	Ethyl methacrylate	0.370	0.417	-12.7	82	0.00
57 T	1,3-Dichloropropane	0.479	0.517	-7.9	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.164	0.195	-18.9	78	0.00
59 T	2-Hexanone	0.176	0.205	-16.5	86	0.00
60 T	Dibromochloromethane	0.332	0.370	-11.4	85	0.00
61 T	1,2-Dibromoethane	0.267	0.287	-7.5	83	0.00
62 S	4-Bromofluorobenzene	0.418	0.330	21.1	69	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
64 T	Tetrachloroethene	0.389	0.386	0.8	79	0.00
65 PM	Chlorobenzene	1.027	1.033	-0.6	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.403	-6.1	84	0.00
67 C	Ethyl Benzene	1.816	1.831	-0.8#	79	0.00
68 T	m/p-Xylenes	0.699	0.706	-1.0	79	0.00
69 T	o-Xylene	0.663	0.679	-2.4	79	0.00
70 T	Styrene	1.124	1.190	-5.9	81	0.00
71 P	Bromoform	0.234	0.257	-9.8	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.633	3.542	2.5	80	0.00
74 T	N-amyl acetate	0.922	0.941	-2.1	83	0.00
75 P	1,1,2,2-Tetrachloroethane	0.762	0.786	-3.1	88	0.00
76 T	1,2,3-Trichloropropane	0.568	0.642	-13.0	96	0.00
77 T	Bromobenzene	0.849	0.832	2.0	82	0.00
78 T	n-propylbenzene	4.420	4.296	2.8	80	0.00
79 T	2-Chlorotoluene	2.489	2.459	1.2	82	0.00
80 T	1,3,5-Trimethylbenzene	3.031	2.982	1.6	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.264	0.273	-3.4	84	0.00
82 T	4-Chlorotoluene	2.592	2.514	3.0	81	0.00
83 T	tert-Butylbenzene	2.621	2.584	1.4	80	0.00
84 T	1,2,4-Trimethylbenzene	2.989	2.988	0.0	82	0.00
85 T	sec-Butylbenzene	3.938	3.911	0.7	81	0.00
86 T	p-Isopropyltoluene	3.267	3.245	0.7	81	0.00
87 T	1,3-Dichlorobenzene	1.695	1.674	1.2	83	0.00
88 T	1,4-Dichlorobenzene	1.685	1.653	1.9	82	0.00
89 T	n-Butylbenzene	3.053	3.073	-0.7	81	0.00
90 T	Hexachloroethane	0.645	0.650	-0.8	84	0.00
91 T	1,2-Dichlorobenzene	1.516	1.527	-0.7	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.126	0.131	-4.0	89	0.00
93 T	1,2,4-Trichlorobenzene	0.906	0.925	-2.1	84	0.00
94 T	Hexachlorobutadiene	0.533	0.544	-2.1	85	0.00
95 T	Naphthalene	1.845	2.029	-10.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102522\
Data File : VY011104.D
Acq On : 25 Oct 2022 19:05
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Oct 26 09:02:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.791	0.839	-6.1	87	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6