

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011253.D
 Acq On : 02 Nov 2022 11:40
 Operator : KP/MD
 Sample : N5361-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-2-102822

Quant Time: Nov 02 13:27:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	340748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	574796	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	492927	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	212879	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	170453	46.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.540%
35) Dibromofluoromethane	7.716	113	177436	48.046	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.100%
50) Toluene-d8	10.179	98	629587	43.909	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.820%
62) 4-Bromofluorobenzene	12.477	95	209801	42.147	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011253.D
Acq On : 02 Nov 2022 11:40
Operator : KP/MD
Sample : N5361-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-2-102822

Quant Time: Nov 02 13:27:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

