

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011256.D
 Acq On : 02 Nov 2022 12:48
 Operator : KP/MD
 Sample : N5337-05
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-03-14.5-15.0

Quant Time: Nov 03 05:14:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	30018	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	65729	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	85509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	43978	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	39171	122.011	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 244.020%#		
35) Dibromofluoromethane	7.715	113	32530	77.029	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 154.060%#		
50) Toluene-d8	10.178	98	91584	55.856	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 111.720%		
62) 4-Bromofluorobenzene	12.477	95	47185	82.894	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 165.780%#		
Target Compounds						Qvalue
16) Acetone	3.936	43	1744	17.996	ug/l #	75
20) Methylene Chloride	4.710	84	6047	11.373	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011256.D
Acq On : 02 Nov 2022 12:48
Operator : KP/MD
Sample : N5337-05
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-03-14.5-15.0

Quant Time: Nov 03 05:14:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

