

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011261.D
 Acq On : 02 Nov 2022 14:41
 Operator : KP/MD
 Sample : N5395-04
 Misc : 5.82g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-10-103122

Quant Time: Nov 03 05:16:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	229118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	351196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	220052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	57140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	96304	39.301	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.600%
35) Dibromofluoromethane	7.716	113	117794	52.204	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.400%
50) Toluene-d8	10.179	98	393198	44.882	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.760%
62) 4-Bromofluorobenzene	12.477	95	76659	25.205	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	50.420%
Target Compounds						
16) Acetone	3.948	43	8937	12.082	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011261.D
Acq On : 02 Nov 2022 14:41
Operator : KP/MD
Sample : N5395-04
Misc : 5.82g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-10-103122

Quant Time: Nov 03 05:16:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

