

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011265.D
 Acq On : 02 Nov 2022 16:12
 Operator : KP/MD
 Sample : N5395-09
 Misc : 5.74g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-6-103122

Quant Time: Nov 03 05:17:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	247853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	413181	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	297668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	90663	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	148859	56.156	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	112.320%	
35) Dibromofluoromethane	7.716	113	159915	60.239	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	120.480%	
50) Toluene-d8	10.179	98	517803	50.238	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	100.480%	
62) 4-Bromofluorobenzene	12.477	95	132292	36.972	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	73.940%	
Target Compounds						
16) Acetone	3.948	43	2566	3.207	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011265.D
Acq On : 02 Nov 2022 16:12
Operator : KP/MD
Sample : N5395-09
Misc : 5.74g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-6-103122

Quant Time: Nov 03 05:17:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

