

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011268.D
 Acq On : 02 Nov 2022 17:20
 Operator : KP/MD
 Sample : N5395-12
 Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-9-103122

Quant Time: Nov 03 05:18:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	234067	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	363415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	237668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	60438	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	114066	45.565	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.140%
35) Dibromofluoromethane	7.715	113	134342	57.536	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.080%
50) Toluene-d8	10.178	98	439865	48.520	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.040%
62) 4-Bromofluorobenzene	12.477	95	94395	29.993	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	59.980%
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	25511	33.760	ug/l	90
25) 2-Butanone	6.990	43	7613	7.160	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011268.D
Acq On : 02 Nov 2022 17:20
Operator : KP/MD
Sample : N5395-12
Misc : 5.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-9-103122

Quant Time: Nov 03 05:18:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

