

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011270.D
 Acq On : 02 Nov 2022 18:05
 Operator : KP/MD
 Sample : N5336-02
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-16.5-17.0RE

Quant Time: Nov 03 05:19:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	4620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	7990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	6817	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	1813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	3460	70.025	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	140.040%
35) Dibromofluoromethane	7.716	113	1515	29.512	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	59.020%
50) Toluene-d8	10.179	98	8898	44.643	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.280%
62) 4-Bromofluorobenzene	12.483	95	2662	38.471	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.940%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011270.D
Acq On : 02 Nov 2022 18:05
Operator : KP/MD
Sample : N5336-02
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-16.5-17.0RE

Quant Time: Nov 03 05:19:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

