

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011271.D
 Acq On : 02 Nov 2022 18:28
 Operator : KP/MD
 Sample : N5336-03
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-01-16.5-17.0-DUP

Quant Time: Nov 03 05:19:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	307502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	534464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	479347	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	215697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	169601	51.570	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.140%	
35) Dibromofluoromethane	7.715	113	175602	51.138	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.280%	
50) Toluene-d8	10.178	98	623927	46.798	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.600%	
62) 4-Bromofluorobenzene	12.477	95	214593	46.363	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.720%	
Target Compounds						
16) Acetone	3.948	43	7153	7.205	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011271.D
Acq On : 02 Nov 2022 18:28
Operator : KP/MD
Sample : N5336-03
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-01-16.5-17.0-DUP

Quant Time: Nov 03 05:19:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

