

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011249.D
 Acq On : 02 Nov 2022 09:42
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/03/2022
 Supervised By :Mahesh Dadoda 11/03/2022

Quant Time: Nov 02 13:25:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	243488	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	386181	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	353439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	179599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121024	46.474	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	92.940%		
35) Dibromofluoromethane	7.709	113	124157	50.039	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	100.080%		
50) Toluene-d8	10.178	98	470992	48.891	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	97.780%		
62) 4-Bromofluorobenzene	12.477	95	169653	50.728	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	101.460%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.906	85	92885	48.108	ug/l	99
3) Chloromethane	2.113	50	128742	47.683	ug/l	100
4) Vinyl Chloride	2.247	62	143862	47.700	ug/l	100
5) Bromomethane	2.643	94	99203	49.304	ug/l	99
6) Chloroethane	2.790	64	94442	48.731	ug/l	99
7) Trichlorofluoromethane	3.125	101	216981	48.698	ug/l	94
8) Diethyl Ether	3.527	74	78661	49.201	ug/l	82
9) 1,1,2-Trichlorotrifluo...	3.893	101	133199	49.723	ug/l	94
10) Methyl Iodide	4.088	142	174585	50.629	ug/l	91
11) Tert butyl alcohol	4.948	59	63746	237.320	ug/l #	95
12) 1,1-Dichloroethene	3.869	96	133604	50.444	ug/l	83
13) Acrolein	3.722	56	47078	345.322	ug/l	98
14) Allyl chloride	4.472	41	200226	49.908	ug/l #	85
15) Acrylonitrile	5.155	53	200644	250.102	ug/l	97
16) Acetone	3.942	43	213434	271.516	ug/l #	82
17) Carbon Disulfide	4.192	76	424848	48.838	ug/l	100
18) Methyl Acetate	4.472	43	95215	45.752	ug/l #	87
19) Methyl tert-butyl Ether	5.216	73	360241	51.184	ug/l	96
20) Methylene Chloride	4.710	84	155627	48.232	ug/l #	83
21) trans-1,2-Dichloroethene	5.210	96	153795	51.053	ug/l	87
22) Diisopropyl ether	6.112	45	426569	51.283	ug/l #	92
23) Vinyl Acetate	6.051	43	1323757	260.740	ug/l #	91
24) 1,1-Dichloroethane	6.008	63	261794	50.257	ug/l	98
25) 2-Butanone	6.978	43	272249	246.145	ug/l #	85
26) 2,2-Dichloropropane	6.978	77	231325	48.566	ug/l	94
27) cis-1,2-Dichloroethene	6.978	96	171320	51.613	ug/l	87
28) Bromochloromethane	7.331	49	102358	51.156	ug/l #	82
29) Tetrahydrofuran	7.344	42	161722	251.160	ug/l #	83
30) Chloroform	7.502	83	267565	50.858	ug/l	100
31) Cyclohexane	7.783	56	232096	48.086	ug/l	92
32) 1,1,1-Trichloroethane	7.697	97	236543	50.691	ug/l	95
36) 1,1-Dichloropropene	7.917	75	209592	51.726	ug/l	97
37) Ethyl Acetate	7.069	43	107931	49.413	ug/l #	92
38) Carbon Tetrachloride	7.898	117	217251	51.681	ug/l	98
39) Methylcyclohexane	9.179	83	258801	52.500	ug/l	92
40) Benzene	8.154	78	611447	51.345	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
 Data File : VY011249.D
 Acq On : 02 Nov 2022 09:42
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/03/2022
 Supervised By :Mahesh Dadoda 11/03/2022

Quant Time: Nov 02 13:25:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	62212m	52.418	ug/l	
42) 1,2-Dichloroethane	8.234	62	165362	51.136	ug/l	91
43) Isopropyl Acetate	8.270	43	191936	49.875	ug/l #	87
44) Trichloroethene	8.935	130	165223	51.843	ug/l	96
45) 1,2-Dichloropropane	9.209	63	152160	51.683	ug/l	96
46) Dibromomethane	9.301	93	86245	51.847	ug/l	97
47) Bromodichloromethane	9.490	83	205843	52.201	ug/l	100
48) Methyl methacrylate	9.288	41	91872	51.975	ug/l #	78
49) 1,4-Dioxane	9.294	88	24423	1091.362	ug/l #	85
51) 4-Methyl-2-Pentanone	10.069	43	543308	259.648	ug/l #	86
52) Toluene	10.239	92	389267	52.717	ug/l	96
53) t-1,3-Dichloropropene	10.465	75	215336	51.852	ug/l	100
54) cis-1,3-Dichloropropene	9.922	75	248463	52.287	ug/l #	81
55) 1,1,2-Trichloroethane	10.642	97	122831	51.521	ug/l	99
56) Ethyl methacrylate	10.508	69	169873	53.943	ug/l #	74
57) 1,3-Dichloropropane	10.788	76	210420	51.563	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.782	63	384960	260.752	ug/l	96
59) 2-Hexanone	10.831	43	400726	268.246	ug/l	83
60) Dibromochloromethane	10.983	129	147519	52.079	ug/l	99
61) 1,2-Dibromoethane	11.087	107	117639	51.974	ug/l	99
64) Tetrachloroethene	10.715	164	159169	53.017	ug/l	96
65) Chlorobenzene	11.514	112	412192	52.861	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.587	131	151382	52.726	ug/l	98
67) Ethyl Benzene	11.593	91	740286	53.520	ug/l	99
68) m/p-Xylenes	11.703	106	578061	107.661	ug/l	95
69) o-Xylene	12.026	106	273375	54.316	ug/l	94
70) Styrene	12.044	104	467736	54.820	ug/l	95
71) Bromoform	12.209	173	93957	52.651	ug/l #	100
73) Isopropylbenzene	12.331	105	714187	53.186	ug/l	100
74) N-amyl acetate	12.142	43	173767	51.552	ug/l #	83
75) 1,1,2,2-Tetrachloroethane	12.580	83	140985	50.738	ug/l	99
76) 1,2,3-Trichloropropane	12.629	75	106675m	48.789	ug/l	
77) Bromobenzene	12.605	156	167064	52.892	ug/l	98
78) n-propylbenzene	12.672	91	867516	52.797	ug/l	98
79) 2-Chlorotoluene	12.751	91	491678	52.225	ug/l	98
80) 1,3,5-Trimethylbenzene	12.812	105	607416	53.457	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.373	75	51498	50.407	ug/l #	83
82) 4-Chlorotoluene	12.855	91	511794	52.752	ug/l	99
83) tert-Butylbenzene	13.074	119	517906	52.897	ug/l	94
84) 1,2,4-Trimethylbenzene	13.117	105	602141	53.580	ug/l	99
85) sec-Butylbenzene	13.251	105	775241	53.405	ug/l	100
86) p-Isopropyltoluene	13.367	119	648989	53.638	ug/l	98
87) 1,3-Dichlorobenzene	13.367	146	330645	52.206	ug/l	99
88) 1,4-Dichlorobenzene	13.446	146	330737	52.068	ug/l	98
89) n-Butylbenzene	13.696	91	600405	53.140	ug/l	99
90) Hexachloroethane	13.958	117	124210	51.331	ug/l	91
91) 1,2-Dichlorobenzene	13.739	146	297280	52.548	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	23656	49.697	ug/l	82
93) 1,2,4-Trichlorobenzene	15.007	180	182605	53.404	ug/l	99
94) Hexachlorobutadiene	15.110	225	103418	51.944	ug/l	99
95) Naphthalene	15.232	128	385610	48.372	ug/l	98
96) 1,2,3-Trichlorobenzene	15.421	180	159611	53.116	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110222\
Data File : VY011249.D
Acq On : 02 Nov 2022 09:42
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022

Quant Time: Nov 02 13:25:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:168 Resp: 24348
Ion Ratio Lower Upper
168 100
99 52.6 46.9 70.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#2
Dichlorodifluoromethane
Concen: 48.108 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Tgt Ion: 85 Resp: 92885
Ion Ratio Lower Upper
85 100
87 32.5 16.1 48.2





#3
Chloromethane
Concen: 47.683 ug/l
RT: 2.113 min Scan# 48
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 50 Resp: 12874
Ion Ratio Lower Upper
50 100
52 30.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#4
Vinyl Chloride
Concen: 47.700 ug/l
RT: 2.247 min Scan# 70
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion: 62 Resp: 143862
Ion Ratio Lower Upper
62 100
64 31.3 25.0 37.6









#5
Bromomethane
Concen: 49.304 ug/l
RT: 2.643 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 93.3 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#6
Chloroethane
Concen: 48.731 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 64 Resp: 94442
Ion Ratio Lower Upper
64 100
66 31.1 25.4 38.0





#7
Trichlorofluoromethane
Concen: 48.698 ug/l
RT: 3.125 min Scan# 213
Delta R.T. 0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 21698
Ion Ratio Lower Upper
101 100
103 64.4 48.0 72.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#8
Diethyl Ether
Concen: 49.201 ug/l
RT: 3.527 min Scan# 280
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion: 74 Resp: 78661
Ion Ratio Lower Upper
74 100
45 91.0 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.723 ug/l
RT: 3.893 min Scan# 340
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 133195
Ion Ratio Lower Upper
101 100
85 42.4 40.6 60.8
151 78.7 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#10
Methyl Iodide
Concen: 50.629 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion:142 Resp: 174585
Ion Ratio Lower Upper
142 100
127 43.7 41.5 62.3
141 14.4 11.9 17.9





#11
Tert butyl alcohol
Concen: 237.320 ug/l
RT: 4.948 min Scan# 514
Delta R.T. -0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 59 Resp: 63740
Ion Ratio Lower Upper
59 100
57 9.6 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#12
1,1-Dichloroethene
Concen: 50.444 ug/l
RT: 3.869 min Scan# 336
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 96 Resp: 133604
Ion Ratio Lower Upper
96 100
61 150.7 145.8 218.6
98 62.1 52.5 78.7





#13
Acrolein
Concen: 345.322 ug/l
RT: 3.722 min Scan# 312
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 47078
Ion Ratio Lower Upper
56 100
55 70.1 57.6 86.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#14
Allyl chloride
Concen: 49.908 ug/l
RT: 4.472 min Scan# 435
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 41 Resp: 200226
Ion Ratio Lower Upper
41 100
39 62.1 58.8 88.2
76 42.1 25.1 37.7#





#15
Acrylonitrile
Concen: 250.102 ug/l
RT: 5.155 min Scan# 547
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 53 Resp: 200644
Ion Ratio Lower Upper
53 100
52 80.8 67.0 100.4
51 35.6 29.8 44.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#16
Acetone
Concen: 271.516 ug/l
RT: 3.942 min Scan# 348
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion: 43 Resp: 213434
Ion Ratio Lower Upper
43 100
58 36.9 22.0 33.0#









#17
Carbon Disulfide
Concen: 48.838 ug/l
RT: 4.192 min Scan# 389
Delta R.T. 0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 76 Resp: 42484
Ion Ratio Lower Upper
76 100
78 9.2 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#18
Methyl Acetate
Concen: 45.752 ug/l
RT: 4.472 min Scan# 435
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 43 Resp: 95215
Ion Ratio Lower Upper
43 100
74 26.5 16.6 24.8#





#19
Methyl tert-butyl Ether
Concen: 51.184 ug/l
RT: 5.216 min Scan# 511
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 73 Resp: 36024
Ion Ratio Lower Upper
73 100
57 21.4 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#20
Methylene Chloride
Concen: 48.232 ug/l
RT: 4.710 min Scan# 474
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 84 Resp: 155627
Ion Ratio Lower Upper
84 100
49 112.8 114.4 171.6#
51 34.0 35.3 52.9#
86 64.5 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 51.053 ug/l
RT: 5.210 min Scan# 511
Delta R.T. -0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 96 Resp: 15379
Ion Ratio Lower Upper
96 100
61 127.6 120.5 180.7
98 63.4 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#22
Diisopropyl ether
Concen: 51.283 ug/l
RT: 6.112 min Scan# 704
Delta R.T. 0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Tgt Ion: 45 Resp: 426569
Ion Ratio Lower Upper
45 100
43 56.0 49.3 73.9
87 29.8 19.2 28.8#
59 12.5 8.4 12.6





#23
 Vinyl Acetate
 Concen: 260.740 ug/l
 RT: 6.051 min Scan# 694
 Delta R.T. -0.007 min
 Lab File: VY011249.D
 Acq: 02 Nov 2022 09:42




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 43 Resp: 132375
 Ion Ratio Lower Upper
 43 100
 86 12.3 7.1 10.7




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 11/03/2022
 Supervised By :Mahesh Dadoda 11/03/2022



#24
 1,1-Dichloroethane
 Concen: 50.257 ug/l
 RT: 6.008 min Scan# 687
 Delta R.T. -0.007 min
 Lab File: VY011249.D
 Acq: 02 Nov 2022 09:42


Tgt Ion: 63 Resp: 261794
 Ion Ratio Lower Upper
 63 100
 98 7.7 3.6 10.8
 100 4.5 2.0 5.8





#25
2-Butanone
Concen: 246.145 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 272249
Ion Ratio Lower Upper
43 100
72 29.3 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#26
2,2-Dichloropropane
Concen: 48.566 ug/l
RT: 6.978 min Scan# 846
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Tgt Ion: 77 Resp: 231325
Ion Ratio Lower Upper
77 100
97 23.7 10.4 31.4





#27
 cis-1,2-Dichloroethene
 Concen: 51.613 ug/l
 RT: 6.978 min Scan# 846
 Delta R.T. -0.000 min
 Lab File: VY011249.D
 Acq: 02 Nov 2022 09:42




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050



Tgt Ion: 96 Resp: 171320
 Ion Ratio Lower Upper
 96 100
 61 136.5 0.0 319.8
 98 64.2 0.0 127.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 11/03/2022
 Supervised By :Mahesh Dadoda 11/03/2022



#28
 Bromochloromethane
 Concen: 51.156 ug/l
 RT: 7.331 min Scan# 904
 Delta R.T. -0.001 min
 Lab File: VY011249.D
 Acq: 02 Nov 2022 09:42

Tgt Ion: 49 Resp: 102358
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 2.2
 130 84.0 55.4 83.0#





#29

Tetrahydrofuran

Concen: 251.160 ug/l

RT: 7.344 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY011249.D


Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050






Tgt Ion:	42	Resp:	16172
Ion Ratio	Lower	Upper	
42	100		
72	48.9	30.3	45.5
71	45.5	29.0	43.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#30

Chloroform

Concen: 50.858 ug/l




RT: 7.502 min Scan# 932


Delta R.T. 0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42


Tgt Ion:	83	Resp:	267565
Ion Ratio	Lower	Upper	
83	100		
85	64.7	51.8	77.6





#31
Cyclohexane
Concen: 48.086 ug/l
RT: 7.783 min Scan# 911
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 232090
Ion Ratio Lower Upper
56 100
69 34.7 26.7 40.1
84 92.9 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#32
1,1,1-Trichloroethane
Concen: 50.691 ug/l
RT: 7.697 min Scan# 964
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion: 97 Resp: 236543
Ion Ratio Lower Upper
97 100
99 64.4 50.7 76.1
61 40.1 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 46.474 ug/l

RT: 8.136 min Scan# 1036

Delta R.T. 0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 121024

Ion Ratio Lower Upper

65 100





67 54.9 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42




Tgt Ion:114 Resp: 386181


Ion Ratio Lower Upper

114 100

63 19.3 0.0 43.2

88 15.4 0.0 31.8





#35

Dibromofluoromethane

Concen: 50.039 ug/l

RT: 7.709 min Scan# 9

Delta R.T. -0.001 min

Lab File: VY011249.D


Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050






Tgt Ion:	113	Resp:	12415
Ion Ratio	Lower	Upper	
113	100		
111	103.7	82.6	124.0
192	19.7	17.6	26.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#36

1,1-Dichloropropene

Concen: 51.726 ug/l




RT: 7.917 min Scan# 1000


Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42


Tgt Ion:	75	Resp:	209592
Ion Ratio	Lower	Upper	
75	100		
110	36.4	17.1	51.2
77	31.3	24.4	36.6





#37
Ethyl Acetate
Concen: 49.413 ug/l
RT: 7.069 min Scan# 861
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 10793
Ion Ratio Lower Upper
43 100
61 15.2 9.7 14.5
70 11.8 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#38
Carbon Tetrachloride
Concen: 51.681 ug/l
RT: 7.898 min Scan# 997
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion:117 Resp: 217251
Ion Ratio Lower Upper
117 100
119 94.5 77.9 116.9
121 30.7 24.8 37.2





#39
Methylcyclohexane
Concen: 52.500 ug/l
RT: 9.179 min Scan# 1207
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 83 Resp: 25880
Ion Ratio Lower Upper
83 100
55 71.8 66.5 99.7
98 48.0 37.9 56.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#40
Benzene
Concen: 51.345 ug/l
RT: 8.154 min Scan# 1039
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 78 Resp: 611447
Ion Ratio Lower Upper
78 100
77 23.1 20.4 30.6





#41
Methacrylonitrile
Concen: 52.418 ug/l m
RT: 7.319 min Scan# 900
Delta R.T. 0.012 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 62212
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#42
1,2-Dichloroethane
Concen: 51.136 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion: 62 Resp: 165362
Ion Ratio Lower Upper
62 100
98 10.4 0.0 14.8





#43

Isopropyl Acetate

Concen: 49.875 ug/l

RT: 8.270 min Scan# 1058

Delta R.T. -0.001 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	43	Resp:	191930
Ion	Ratio	Lower	Upper
43	100		
61	32.1	20.2	30.2
87	15.4	8.9	13.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#44

Trichloroethene

Concen: 51.843 ug/l




RT: 8.935 min Scan# 1167


Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Tgt Ion:	130	Resp:	165223
Ion	Ratio	Lower	Upper
130	100		
95	97.7	0.0	188.2





#45

1,2-Dichloropropane

Concen: 51.683 ug/l

RT: 9.209 min Scan# 1212

Delta R.T. -0.007 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 152160

Ion Ratio Lower Upper

63 100





65 30.9 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#46

Dibromomethane

Concen: 51.847 ug/l

RT: 9.301 min Scan# 1227

Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42




Tgt Ion: 93 Resp: 86245


Ion Ratio Lower Upper

93 100

95 82.6 64.3 96.5

174 100.6 83.4 125.0









#47
Bromodichloromethane
Concen: 52.201 ug/l
RT: 9.490 min Scan# 1258
Delta R.T. -0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 83 Resp: 20584
Ion Ratio Lower Upper
83 100
85 64.6 51.9 77.9
127 8.5 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#48
Methyl methacrylate
Concen: 51.975 ug/l
RT: 9.288 min Scan# 1225
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion: 41 Resp: 91872
Ion Ratio Lower Upper
41 100
69 97.0 58.2 87.2#
39 48.8 47.4 71.0





#49

1,4-Dioxane

Concen: 1091.362 ug/l

RT: 9.294 min Scan# 1226

Delta R.T. -0.001 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	88	Resp:	2442
Ion	Ratio	Lower	Upper
88	100		
43	28.6	32.8	49.2
58	64.1	59.8	89.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#50

Toluene-d8

Concen: 48.891 ug/l




RT: 10.178 min Scan# 1371


Delta R.T. -0.001 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42


Tgt Ion:	98	Resp:	470992
Ion	Ratio	Lower	Upper
98	100		
100	65.5	50.2	75.2





#51
4-Methyl-2-Pentanone
Concen: 259.648 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 543308
Ion Ratio Lower Upper
43 100
58 43.2 28.1 42.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#52
Toluene
Concen: 52.717 ug/l
RT: 10.239 min Scan# 1381
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion: 92 Resp: 389267
Ion Ratio Lower Upper
92 100
91 172.1 142.6 213.8









#53
t-1,3-Dichloropropene
Concen: 51.852 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050

Tgt Ion: 75 Resp: 215330
Ion Ratio Lower Upper
75 100
77 32.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#54
cis-1,3-Dichloropropene
Concen: 52.287 ug/l
RT: 9.922 min Scan# 1329
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 75 Resp: 248463
Ion Ratio Lower Upper
75 100
77 31.4 25.6 38.4
39 38.5 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 51.521 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.8	69.8	104.8
85	54.6	44.3	66.5
99	62.1	49.8	74.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#56
Ethyl methacrylate
Concen: 53.943 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion	Ratio	Lower	Upper
69	100		
41	57.3	63.4	95.0#
39	26.1	35.8	53.8#





#57

1,3-Dichloropropane

Concen: 51.563 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 210420

Ion Ratio Lower Upper

76 100





78 32.2 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#58

2-Chloroethyl Vinyl ether

Concen: 260.752 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. -0.001 min

Lab File: VY011249.D




Acq: 02 Nov 2022 09:42


Tgt Ion: 63 Resp: 384960

Ion Ratio Lower Upper

63 100


106 28.9 21.5 32.3





#59
2-Hexanone
Concen: 268.246 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 400720
Ion Ratio Lower Upper
43 100
58 60.6 24.4 73.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#60
Dibromochloromethane
Concen: 52.079 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Tgt Ion:129 Resp: 147519
Ion Ratio Lower Upper
129 100
127 77.9 38.7 116.1





#61
1,2-Dibromoethane
Concen: 51.974 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 117639
Ion Ratio Lower Upper
107 100
109 95.1 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#62
4-Bromofluorobenzene
Concen: 50.728 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Tgt Ion: 95 Resp: 169653
Ion Ratio Lower Upper
95 100
174 81.9 0.0 177.0
176 79.3 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 117
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 353439
Ion Ratio Lower Upper
117 100
82 53.4 43.0 64.4
119 32.6 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#64
Tetrachloroethene
Concen: 53.017 ug/l
RT: 10.715 min Scan# 1459
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion:164 Resp: 159169
Ion Ratio Lower Upper
164 100
166 128.4 100.4 150.6
129 94.9 71.4 107.0
131 92.2 70.8 106.2









#65
Chlorobenzene
Concen: 52.861 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:112 Resp: 412192
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 52.726 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion:131 Resp: 151382
Ion Ratio Lower Upper
131 100
133 95.1 47.9 143.6
119 65.1 30.8 92.3





#67
Ethyl Benzene
Concen: 53.520 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 740280
Ion Ratio Lower Upper
91 100
106 31.3 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#68
m/p-Xylenes
Concen: 107.661 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion:106 Resp: 578061
Ion Ratio Lower Upper
106 100
91 200.8 166.6 249.8









#69
o-Xylene
Concen: 54.316 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 273375
Ion Ratio Lower Upper
106 100
91 210.3 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#70
Styrene
Concen: 54.820 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion:104 Resp: 467736
Ion Ratio Lower Upper
104 100
78 47.6 42.4 63.6
103 54.1 44.4 66.6





#71

Bromoform

Concen: 52.651 ug/l

RT: 12.209 min Scan# 1704

Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	173	Resp:	9395
Ion Ratio	Lower	Upper	
173	100		
175	49.2	24.6	74.0
254	0.1	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 13.422 min Scan# 1903


Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Tgt Ion:	152	Resp:	179599
Ion Ratio	Lower	Upper	
152	100		
115	58.5	28.2	84.7
150	174.8	0.0	347.6









#73
Isopropylbenzene
Concen: 53.186 ug/l
RT: 12.331 min Scan# 1724
Delta R.T. 0.006 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:105 Resp: 71418
Ion Ratio Lower Upper
105 100
120 26.6 13.4 40.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#74
N-ethyl acetate
Concen: 51.552 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 43 Resp: 173767
Ion Ratio Lower Upper
43 100
70 48.7 28.7 43.1#
55 27.4 17.3 25.9#
61 27.3 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 50.738 ug/l
RT: 12.580 min Scan# 1765
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 140985
Ion Ratio Lower Upper
83 100
131 10.5 5.6 16.8
85 64.1 32.3 96.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#76
1,2,3-Trichloropropane
Concen: 48.789 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion: 75 Resp: 106675
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 52.892 ug/l

RT: 12.605 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:	156	Resp:	167064
Ion Ratio	Lower	Upper	
156	100		
77	187.9	95.5	286.4
158	97.7	48.0	144.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#78

n-propylbenzene

Concen: 52.797 ug/l




RT: 12.672 min Scan# 1780


Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Tgt Ion:	91	Resp:	867516
Ion Ratio	Lower	Upper	
91	100		
120	23.4	11.3	33.9









#79
2-Chlorotoluene
Concen: 52.225 ug/l
RT: 12.751 min Scan# 1793
Delta R.T. -0.007 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 491678
Ion Ratio Lower Upper
91 100
126 34.8 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#80
1,3,5-Trimethylbenzene
Concen: 53.457 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion:105 Resp: 607416
Ion Ratio Lower Upper
105 100
120 49.3 24.4 73.4









#81
trans-1,4-Dichloro-2-butene
Concen: 50.407 ug/l
RT: 12.373 min Scan# 1731
Delta R.T. -0.007 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 51498
Ion Ratio Lower Upper
75 100
53 80.9 85.3 127.9
89 46.0 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#82
4-Chlorotoluene
Concen: 52.752 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42


Tgt Ion: 91 Resp: 511794
Ion Ratio Lower Upper
91 100
126 34.4 17.0 51.0





#83
tert-Butylbenzene
Concen: 52.897 ug/l
RT: 13.074 min Scan# 1846
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:119 Resp: 517900
Ion Ratio Lower Upper
119 100
91 62.8 34.8 104.4
134 26.2 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#84
1,2,4-Trimethylbenzene
Concen: 53.580 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion:105 Resp: 602141
Ion Ratio Lower Upper
105 100
120 45.3 22.3 66.9





#85

sec-Butylbenzene

Concen: 53.405 ug/l

RT: 13.251 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 77524

Ion Ratio Lower Upper

105 100





134 20.5 10.2 30.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#86

p-Isopropyltoluene

Concen: 53.638 ug/l

RT: 13.367 min Scan# 1894

Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42




Tgt Ion:119 Resp: 648989


Ion Ratio Lower Upper

119 100

134 26.6 13.2 39.5


91 22.8 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 52.206 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 330649
Ion Ratio Lower Upper
146 100
111 40.1 19.5 58.5
148 63.8 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#88
1,4-Dichlorobenzene
Concen: 52.068 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion:146 Resp: 330737
Ion Ratio Lower Upper
146 100
111 38.2 19.3 57.8
148 62.7 32.6 97.8





#89

n-Butylbenzene

Concen: 53.140 ug/l

RT: 13.696 min Scan# 1948

Delta R.T. 0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 91	Resp: 60040
Ion Ratio	Lower Upper
91	100
92	53.6 26.3 78.9
134	26.6 13.0 38.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#90

Hexachloroethane

Concen: 51.331 ug/l




RT: 13.958 min Scan# 1991


Delta R.T. -0.001 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42


Tgt Ion:117	Resp: 124210
Ion Ratio	Lower Upper
117	100
201	78.0 43.0 128.8





#91
1,2-Dichlorobenzene
Concen: 52.548 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 297280
Ion Ratio Lower Upper
146 100
111 41.4 20.3 60.9
148 63.7 32.3 96.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 49.697 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42



Tgt Ion: 75 Resp: 23656
Ion Ratio Lower Upper
75 100
155 88.9 38.0 113.9
157 114.4 47.8 143.3





#93

1,2,4-Trichlorobenzene

Concen: 53.404 ug/l

RT: 15.007 min Scan# 2163

Delta R.T. -0.000 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:180 Resp: 18260

Ion Ratio Lower Upper

180 100

182 94.9 47.3 141.9





145 30.7 14.6 43.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 11/03/2022

Supervised By :Mahesh Dadoda 11/03/2022



#94

Hexachlorobutadiene

Concen: 51.944 ug/l

RT: 15.110 min Scan# 2180

Delta R.T. -0.001 min

Lab File: VY011249.D

Acq: 02 Nov 2022 09:42




Tgt Ion:225 Resp: 103418


Ion Ratio Lower Upper

225 100

223 62.8 30.8 92.3


227 63.4 31.6 94.7





#95
Naphthalene
Concen: 48.372 ug/l
RT: 15.232 min Scan# 2110
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 385610
Ion Ratio Lower Upper
128 100
127 12.7 11.0 16.4
129 10.9 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/03/2022
Supervised By :Mahesh Dadoda 11/03/2022



#96
1,2,3-Trichlorobenzene
Concen: 53.116 ug/l
RT: 15.421 min Scan# 2231
Delta R.T. -0.001 min
Lab File: VY011249.D
Acq: 02 Nov 2022 09:42

Tgt Ion:180 Resp: 159611
Ion Ratio Lower Upper
180 100
182 95.4 47.3 141.9
145 32.5 15.3 45.9

