

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110322\
 Data File : VY011282.D
 Acq On : 03 Nov 2022 12:50
 Operator : KP/MD
 Sample : N5392-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OE-1-103122

Quant Time: Nov 03 14:30:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	164782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	278241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	244978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	107054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	102967	58.426	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.860%
35) Dibromofluoromethane	7.716	113	103733	58.026	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.060%
50) Toluene-d8	10.179	98	339684	48.940	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.477	95	114174	47.383	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110322\
Data File : VY011282.D
Acq On : 03 Nov 2022 12:50
Operator : KP/MD
Sample : N5392-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OE-1-103122

Quant Time: Nov 03 14:30:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

