

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110322\
 Data File : VY011288.D
 Acq On : 03 Nov 2022 15:39
 Operator : KP/MD
 Sample : N5395-03RE
 Misc : 5.97g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-11-103122RE

Quant Time: Nov 04 02:21:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 01 11:56:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	214106	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	368691	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	274560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	79349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	134840	58.885	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.780%
35) Dibromofluoromethane	7.716	113	133388	56.310	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.620%
50) Toluene-d8	10.179	98	432868	47.065	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.140%
62) 4-Bromofluorobenzene	12.477	95	110149	34.498	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	69.000%
Target Compounds						Qvalue
16) Acetone	3.936	43	9484	13.721	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110322\
Data File : VY011288.D
Acq On : 03 Nov 2022 15:39
Operator : KP/MD
Sample : N5395-03RE
Misc : 5.97g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-11-103122RE

Quant Time: Nov 04 02:21:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y110122S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 01 11:56:18 2022
Response via : Initial Calibration

