

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111324\
 Data File : VY020283.D
 Acq On : 13 Nov 2024 15:15
 Operator : SY/MD
 Sample : P4810-02
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Nov 14 00:58:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 15:23:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	24523	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.615	114	40171	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	19481	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.358	152	3557	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	8623	28.174	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	56.340%
35) Dibromofluoromethane	7.640	113	11499	45.023	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.040%
50) Toluene-d8	10.109	98	39586	38.938	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	77.880%
62) 4-Bromofluorobenzene	12.407	95	6201	18.779	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	37.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY111324\
Data File : VY020283.D
Acq On : 13 Nov 2024 15:15
Operator : SY/MD
Sample : P4810-02
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Nov 14 00:58:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103024S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 15:23:13 2024
Response via : Initial Calibration

