

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121125\
 Data File : VY023945.D
 Acq On : 11 Dec 2025 11:09
 Operator : SY/MD
 Sample : Q3827-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT4262

Quant Time: Dec 12 03:20:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121025S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 11 03:57:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	740700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	1276389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	1087589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	458740	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	352004	53.750	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.500%
35) Dibromofluoromethane	7.640	113	396830	47.561	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.120%
50) Toluene-d8	10.109	98	1462050	47.065	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.140%
62) 4-Bromofluorobenzene	12.407	95	430846	41.531	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY121125\
Data File : VY023945.D
Acq On : 11 Dec 2025 11:09
Operator : SY/MD
Sample : Q3827-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT4262

Quant Time: Dec 12 03:20:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y121025S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 11 03:57:51 2025
Response via : Initial Calibration

