

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
 Data File : VY011969.D
 Acq On : 22 Dec 2022 13:04
 Operator : KP/MD
 Sample : N6163-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-122022

Quant Time: Dec 23 04:31:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	7843	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	14329	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	13900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	5030	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	6529	82.628	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	165.260%#
35) Dibromofluoromethane	7.716	113	3850	44.630	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.260%
50) Toluene-d8	10.179	98	17603	50.295	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.580%
62) 4-Bromofluorobenzene	12.483	95	5000	42.477	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
Data File : VY011969.D
Acq On : 22 Dec 2022 13:04
Operator : KP/MD
Sample : N6163-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-122022

Quant Time: Dec 23 04:31:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

