

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
 Data File : VY011970.D
 Acq On : 22 Dec 2022 13:28
 Operator : KP/MD
 Sample : N6160-11
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 23 04:31:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	193972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	346041	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	324733	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	142159	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	123781	63.340	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.680%
35) Dibromofluoromethane	7.716	113	108771	52.212	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.420%
50) Toluene-d8	10.179	98	422395	49.974	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.940%
62) 4-Bromofluorobenzene	12.483	95	131597	46.293	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
Data File : VY011970.D
Acq On : 22 Dec 2022 13:28
Operator : KP/MD
Sample : N6160-11
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 23 04:31:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

