

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
 Data File : VY011976.D
 Acq On : 22 Dec 2022 15:48
 Operator : KP/MD
 Sample : N6151-07
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FS-02-(0-2)

Quant Time: Dec 23 04:32:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	194013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	347476	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	322829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	142864	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	122384	62.612	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	125.220%	
35) Dibromofluoromethane	7.716	113	110082	52.623	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	105.240%	
50) Toluene-d8	10.179	98	421222	49.629	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	99.260%	
62) 4-Bromofluorobenzene	12.483	95	133596	46.802	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	93.600%	
Target Compounds						
52) Toluene	10.246	92	29524	4.826	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
Data File : VY011976.D
Acq On : 22 Dec 2022 15:48
Operator : KP/MD
Sample : N6151-07
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FS-02-(0-2)

Quant Time: Dec 23 04:32:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

