

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
 Data File : VY011985.D
 Acq On : 22 Dec 2022 19:18
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 04:33:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.377	0.345	8.5	71	0.00
3 P	Chloromethane	0.445	0.376	15.5	72	0.00
4 C	Vinyl Chloride	0.491	0.410	16.5#	71	0.00
5 T	Bromomethane	0.349	0.305	12.6	78	0.00
6 T	Chloroethane	0.321	0.273	15.0	74	0.00
7 T	Trichlorofluoromethane	0.859	0.744	13.4	75	0.00
8 T	Diethyl Ether	0.265	0.233	12.1	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.405	14.4	76	0.00
10 T	Methyl Iodide	0.654	0.575	12.1	72	0.00
11 T	Tert butyl alcohol	0.065	0.033	49.2#	60	0.00
12 CM	1,1-Dichloroethene	0.473	0.415	12.3#	76	0.00
13 T	Acrolein	0.047	0.036	23.4	77	0.00
14 T	Allyl chloride	0.697	0.584	16.2	73	0.00
15 T	Acrylonitrile	0.123	0.105	14.6	73	0.00
16 T	Acetone	0.137	0.097	29.2#	54	0.00
17 T	Carbon Disulfide	1.509	1.262	16.4	72	0.00
18 T	Methyl Acetate	0.348	0.285	18.1	73	0.00
19 T	Methyl tert-butyl Ether	1.214	1.105	9.0	78	0.00
20 T	Methylene Chloride	0.701	0.520	25.8#	78	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.472	12.3	76	0.00
22 T	Diisopropyl ether	1.372	1.275	7.1	79	0.00
23 T	Vinyl Acetate	0.789	0.719	8.9	75	0.00
24 P	1,1-Dichloroethane	0.892	0.807	9.5	80	0.00
25 T	2-Butanone	0.168	0.128	23.8	62	0.00
26 T	2,2-Dichloropropane	0.873	0.734	15.9	75	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.538	7.4	81	0.00
28 T	Bromochloromethane	0.245	0.223	9.0	86	0.00
29 T	Tetrahydrofuran	0.095	0.081	14.7	71	0.00
30 C	Chloroform	0.961	0.882	8.2#	81	0.00
31 T	Cyclohexane	0.848	0.683	19.5	72	0.00
32 T	1,1,1-Trichloroethane	0.914	0.809	11.5	79	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.486	3.6	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.301	0.296	1.7	89	0.00
36 T	1,1-Dichloropropene	0.483	0.409	15.3	75	0.00
37 T	Ethyl Acetate	0.217	0.182	16.1	75	0.00
38 T	Carbon Tetrachloride	0.558	0.480	14.0	77	0.00
39 T	Methylcyclohexane	0.581	0.498	14.3	74	0.00
40 TM	Benzene	1.382	1.214	12.2	79	0.00
41 T	Methacrylonitrile	0.105	0.079	24.8	63	0.00
42 TM	1,2-Dichloroethane	0.398	0.353	11.3	80	0.00
43 T	Isopropyl Acetate	0.396	0.335	15.4	74	0.00
44 TM	Trichloroethene	0.389	0.341	12.3	80	0.00
45 C	1,2-Dichloropropane	0.319	0.285	10.7#	81	0.00
46 T	Dibromomethane	0.184	0.162	12.0	79	0.00
47 T	Bromodichloromethane	0.467	0.418	10.5	80	0.00
48 T	Methyl methacrylate	0.183	0.160	12.6	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
 Data File : VY011985.D
 Acq On : 22 Dec 2022 19:18
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 04:33:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	75	0.00
50 S	Toluene-d8	1.221	1.190	2.5	88	0.00
51 T	4-Methyl-2-Pentanone	0.211	0.181	14.2	73	0.00
52 CM	Toluene	0.880	0.791	10.1#	79	0.00
53 T	t-1,3-Dichloropropene	0.471	0.418	11.3	78	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.483	8.9	80	0.00
55 T	1,1,2-Trichloroethane	0.261	0.234	10.3	82	0.00
56 T	Ethyl methacrylate	0.330	0.300	9.1	77	0.00
57 T	1,3-Dichloropropane	0.436	0.390	10.6	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.127	0.117	7.9	85	0.00
59 T	2-Hexanone	0.155	0.127	18.1	66	0.00
60 T	Dibromochloromethane	0.329	0.299	9.1	79	0.00
61 T	1,2-Dibromoethane	0.244	0.216	11.5	78	0.00
62 S	4-Bromofluorobenzene	0.411	0.402	2.2	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.438	0.381	13.0	78	0.00
65 PM	Chlorobenzene	1.038	0.934	10.0	81	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.356	7.8	82	0.00
67 C	Ethyl Benzene	1.877	1.667	11.2#	78	0.00
68 T	m/p-Xylenes	0.730	0.651	10.8	78	0.00
69 T	o-Xylene	0.679	0.625	8.0	80	0.00
70 T	Styrene	1.140	1.055	7.5	80	0.00
71 P	Bromoform	0.229	0.205	10.5	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.710	3.330	10.2	78	0.00
74 T	N-amyl acetate	0.761	0.669	12.1	75	0.00
75 P	1,1,2,2-Tetrachloroethane	0.619	0.546	11.8	79	0.00
76 T	1,2,3-Trichloropropane	0.622	0.434	30.2#	62	0.00
77 T	Bromobenzene	0.858	0.772	10.0	81	0.00
78 T	n-propylbenzene	4.460	3.976	10.9	78	0.00
79 T	2-Chlorotoluene	2.507	2.242	10.6	79	0.00
80 T	1,3,5-Trimethylbenzene	3.177	2.874	9.5	79	0.00
81 T	trans-1,4-Dichloro-2-butene	0.227	0.198	12.8	76	0.00
82 T	4-Chlorotoluene	2.594	2.331	10.1	79	0.00
83 T	tert-Butylbenzene	2.692	2.433	9.6	78	0.00
84 T	1,2,4-Trimethylbenzene	3.125	2.842	9.1	79	0.00
85 T	sec-Butylbenzene	4.020	3.544	11.8	77	0.00
86 T	p-Isopropyltoluene	3.392	3.052	10.0	78	0.00
87 T	1,3-Dichlorobenzene	1.726	1.524	11.7	79	0.00
88 T	1,4-Dichlorobenzene	1.716	1.501	12.5	79	0.00
89 T	n-Butylbenzene	3.071	2.670	13.1	75	0.00
90 T	Hexachloroethane	0.647	0.574	11.3	80	0.00
91 T	1,2-Dichlorobenzene	1.505	1.343	10.8	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.106	0.091	14.2	75	0.00
93 T	1,2,4-Trichlorobenzene	0.885	0.785	11.3	75	0.00
94 T	Hexachlorobutadiene	0.572	0.496	13.3	77	0.00
95 T	Naphthalene	1.644	1.475	10.3	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY122222\
Data File : VY011985.D
Acq On : 22 Dec 2022 19:18
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Dec 23 04:33:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.687	8.3	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6