

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
 Data File : VY012043.D
 Acq On : 03 Jan 2023 13:48
 Operator : KP/MD
 Sample : N6248-12
 Misc : 11.42g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-20

Quant Time: Jan 04 01:17:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	49915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	111582	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	140754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	70550	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	68098	135.415	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	270.840%#
35) Dibromofluoromethane	7.716	113	46336	68.977	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	137.960%
50) Toluene-d8	10.179	98	146852	53.881	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.760%
62) 4-Bromofluorobenzene	12.483	95	62596	68.289	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	136.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
Data File : VY012043.D
Acq On : 03 Jan 2023 13:48
Operator : KP/MD
Sample : N6248-12
Misc : 11.42g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-20

Quant Time: Jan 04 01:17:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

