

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
 Data File : VY012046.D
 Acq On : 03 Jan 2023 14:58
 Operator : KP/MD
 Sample : 01001-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-C2

Quant Time: Jan 04 01:18:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	849	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.697	114	1598	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	1423	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	745	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	765	89.437	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 178.880%#	
35) Dibromofluoromethane	7.710	113	228	23.700	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 47.400%#	
50) Toluene-d8	10.179	98	1856	47.550	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 95.100%	
62) 4-Bromofluorobenzene	12.477	95	849	64.674	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 129.340%	
Target Compounds						
27) cis-1,2-Dichloroethene	6.990	96	1100	111.432	ug/l	Qvalue 81
44) Trichloroethene	8.935	130	897	72.063	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
Data File : VY012046.D
Acq On : 03 Jan 2023 14:58
Operator : KP/MD
Sample : 01001-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-C2

Quant Time: Jan 04 01:18:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

