

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
 Data File : VY012051.D
 Acq On : 03 Jan 2023 16:55
 Operator : KP/MD
 Sample : N6237-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-122922RE

Quant Time: Jan 04 01:20:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	4769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	4399	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1170	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	2100	72.527	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	145.060%
35) Dibromofluoromethane	7.716	113	672	23.406	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	46.820%#
50) Toluene-d8	10.179	98	5660	48.589	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.180%
62) 4-Bromofluorobenzene	12.477	95	1546	39.462	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
Data File : VY012051.D
Acq On : 03 Jan 2023 16:55
Operator : KP/MD
Sample : N6237-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-122922RE

Quant Time: Jan 04 01:20:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

