



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010323\
Data File : VY012035.D
Acq On : 03 Jan 2023 09:20
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Title : SW846 8260
Last Update : Tue Dec 20 12:37:37 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	12763	PASS
75	95	30	60	53.2	37648	PASS
95	95	100	100	100.0	70736	PASS
96	95	5	9	6.6	4641	PASS
173	174	0.00	2	1.3	814	PASS
174	95	50	100	90.3	63861	PASS
175	174	5	9	7.7	4920	PASS
176	174	95	101	97.1	62029	PASS
177	176	5	9	6.4	3946	PASS