

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010325\
 Data File : VY020767.D
 Acq On : 03 Jan 2025 10:51
 Operator : SY/MD
 Sample : VY0103SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0103SBL01

Quant Time: Jan 03 22:02:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	153017	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.622	114	252510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	209480	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	81586	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	66625	53.158	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.320%
35) Dibromofluoromethane	7.640	113	75966	54.618	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.240%
50) Toluene-d8	10.109	98	292230	52.738	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.480%
62) 4-Bromofluorobenzene	12.414	95	86459	46.775	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	93.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010325\
Data File : VY020767.D
Acq On : 03 Jan 2025 10:51
Operator : SY/MD
Sample : VY0103SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0103SBL01

Quant Time: Jan 03 22:02:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

