

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010325\
 Data File : VY020771.D
 Acq On : 03 Jan 2025 15:51
 Operator : SY/MD
 Sample : Q1009-03RE
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-03-1-03-2025RE

Quant Time: Jan 03 22:05:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 26 16:12:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.72
34) 1,4-Difluorobenzene	8.622	114	213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.426	117	576	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	339	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	46	0.000	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.109	98	373	78.915	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	157.840%#
62) 4-Bromofluorobenzene	12.414	95	273	175.092	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	350.180%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010325\
Data File : VY020771.D
Acq On : 03 Jan 2025 15:51
Operator : SY/MD
Sample : Q1009-03RE
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-03-1-03-2025RE

Quant Time: Jan 03 22:05:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122624S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 26 16:12:28 2024
Response via : Initial Calibration

