

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
 Data File : VY012066.D
 Acq On : 04 Jan 2023 13:13
 Operator : KP/MD
 Sample : N6241-01RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-2-122922RE

Quant Time: Jan 05 00:40:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	5703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	5256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0m	50.000	ug/l	-13.43

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	2533	83.141	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	166.280%#
35) Dibromofluoromethane	7.716	113	811	23.621	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	47.240%#
50) Toluene-d8	10.179	98	6620	47.523	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.040%
62) 4-Bromofluorobenzene	12.483	95	1846	39.403	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
Data File : VY012066.D
Acq On : 04 Jan 2023 13:13
Operator : KP/MD
Sample : N6241-01RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-2-122922RE

Quant Time: Jan 05 00:40:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

