

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
 Data File : VY012067.D
 Acq On : 04 Jan 2023 13:37
 Operator : KP/MD
 Sample : 01008-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-1-010323

Quant Time: Jan 05 00:41:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	18215	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	29585	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	27505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	11933	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	12082	65.838	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.680%
35) Dibromofluoromethane	7.716	113	10130	56.875	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.740%
50) Toluene-d8	10.179	98	34676	47.986	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.980%
62) 4-Bromofluorobenzene	12.483	95	10192	41.936	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
Data File : VY012067.D
Acq On : 04 Jan 2023 13:37
Operator : KP/MD
Sample : 01008-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-1-010323

Quant Time: Jan 05 00:41:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

