

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
 Data File : VY012069.D
 Acq On : 04 Jan 2023 14:24
 Operator : KP/MD
 Sample : 01004-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CONTAINMENT-STONES

Quant Time: Jan 05 00:41:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	178583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	304491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	284354	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	127683	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	102408	56.919	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.840%
35) Dibromofluoromethane	7.716	113	92832	50.641	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.280%
50) Toluene-d8	10.179	98	363391	48.860	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.720%
62) 4-Bromofluorobenzene	12.477	95	115293	46.092	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
Data File : VY012069.D
Acq On : 04 Jan 2023 14:24
Operator : KP/MD
Sample : 01004-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CONTAINMENT-STONES

Quant Time: Jan 05 00:41:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

