

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
 Data File : VY012072.D
 Acq On : 04 Jan 2023 16:42
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:42:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	83	0.00
2 T	Dichlorodifluoromethane	0.377	0.419	-11.1	79	0.00
3 P	Chloromethane	0.445	0.394	11.5	69	0.00
4 C	Vinyl Chloride	0.491	0.452	7.9#	71	0.00
5 T	Bromomethane	0.349	0.323	7.4	76	0.00
6 T	Chloroethane	0.321	0.307	4.4	76	0.00
7 T	Trichlorofluoromethane	0.859	0.921	-7.2	85	-0.01
8 T	Diethyl Ether	0.265	0.297	-12.1	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.504	-6.6	85	0.00
10 T	Methyl Iodide	0.654	0.655	-0.2	75	-0.01
11 T	Tert butyl alcohol	0.065	0.050	23.1	84	0.00
12 CM	1,1-Dichloroethene	0.473	0.482	-1.9#	81	0.00
13 T	Acrolein	0.047	0.031	34.0#	60	0.00
14 T	Allyl chloride	0.697	0.694	0.4	79	0.00
15 T	Acrylonitrile	0.123	0.143	-16.3	91	0.00
16 T	Acetone	0.137	0.148	-8.0	75	0.00
17 T	Carbon Disulfide	1.509	1.241	17.8	65	-0.01
18 T	Methyl Acetate	0.348	0.382	-9.8	89	0.00
19 T	Methyl tert-butyl Ether	1.214	1.472	-21.3	94	-0.01
20 T	Methylene Chloride	0.701	0.687	2.0	93	0.00
21 T	trans-1,2-Dichloroethene	0.538	0.538	0.0	79	-0.01
22 T	Diisopropyl ether	1.372	1.474	-7.4	83	0.00
23 T	Vinyl Acetate	0.789	0.878	-11.3	84	0.00
24 P	1,1-Dichloroethane	0.892	0.944	-5.8	86	0.00
25 T	2-Butanone	0.168	0.180	-7.1	79	0.00
26 T	2,2-Dichloropropane	0.873	0.937	-7.3	87	0.00
27 T	cis-1,2-Dichloroethene	0.581	0.632	-8.8	86	0.00
28 T	Bromochloromethane	0.245	0.191	22.0	66	0.00
29 T	Tetrahydrofuran	0.095	0.106	-11.6	84	0.00
30 C	Chloroform	0.961	1.058	-10.1#	89	0.00
31 T	Cyclohexane	0.848	0.757	10.7	73	0.00
32 T	1,1,1-Trichloroethane	0.914	1.001	-9.5	89	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.529	-5.0	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.301	0.334	-11.0	88	0.00
36 T	1,1-Dichloropropene	0.483	0.493	-2.1	79	0.00
37 T	Ethyl Acetate	0.217	0.243	-12.0	88	0.00
38 T	Carbon Tetrachloride	0.558	0.617	-10.6	86	0.00
39 T	Methylcyclohexane	0.581	0.585	-0.7	76	0.00
40 TM	Benzene	1.382	1.457	-5.4	82	0.00
41 T	Methacrylonitrile	0.105	0.107	-1.9	74	-0.01
42 TM	1,2-Dichloroethane	0.398	0.461	-15.8	91	0.00
43 T	Isopropyl Acetate	0.396	0.460	-16.2	88	0.00
44 TM	Trichloroethene	0.389	0.426	-9.5	87	0.00
45 C	1,2-Dichloropropane	0.319	0.337	-5.6#	83	0.00
46 T	Dibromomethane	0.184	0.209	-13.6	88	0.00
47 T	Bromodichloromethane	0.467	0.539	-15.4	90	0.00
48 T	Methyl methacrylate	0.183	0.222	-21.3	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
 Data File : VY012072.D
 Acq On : 04 Jan 2023 16:42
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 05 00:42:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 20 12:37:37 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	98	0.00
50 S	Toluene-d8	1.221	1.265	-3.6	81	0.00
51 T	4-Methyl-2-Pentanone	0.211	0.249	-18.0	87	0.00
52 CM	Toluene	0.880	0.960	-9.1#	84	0.00
53 T	t-1,3-Dichloropropene	0.471	0.551	-17.0	90	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.608	-14.7	88	0.00
55 T	1,1,2-Trichloroethane	0.261	0.305	-16.9	93	0.00
56 T	Ethyl methacrylate	0.330	0.412	-24.8	92	0.00
57 T	1,3-Dichloropropane	0.436	0.508	-16.5	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.127	0.134	-5.5	85	0.00
59 T	2-Hexanone	0.155	0.180	-16.1	82	0.00
60 T	Dibromochloromethane	0.329	0.401	-21.9	93	0.00
61 T	1,2-Dibromoethane	0.244	0.285	-16.8	90	0.00
62 S	4-Bromofluorobenzene	0.411	0.461	-12.2	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.438	0.497	-13.5	90	0.00
65 PM	Chlorobenzene	1.038	1.128	-8.7	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.386	0.440	-14.0	91	0.00
67 C	Ethyl Benzene	1.877	2.029	-8.1#	85	0.00
68 T	m/p-Xylenes	0.730	0.798	-9.3	86	0.00
69 T	o-Xylene	0.679	0.767	-13.0	88	0.00
70 T	Styrene	1.140	1.306	-14.6	89	0.00
71 P	Bromoform	0.229	0.281	-22.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.710	3.844	-3.6	88	0.00
74 T	N-amyl acetate	0.761	0.841	-10.5	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.619	0.640	-3.4	90	0.00
76 T	1,2,3-Trichloropropane	0.622	0.508	18.3	70	0.00
77 T	Bromobenzene	0.858	0.915	-6.6	93	0.00
78 T	n-propylbenzene	4.460	4.554	-2.1	87	0.00
79 T	2-Chlorotoluene	2.507	2.574	-2.7	88	0.00
80 T	1,3,5-Trimethylbenzene	3.177	3.336	-5.0	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.227	0.247	-8.8	93	0.00
82 T	4-Chlorotoluene	2.594	2.671	-3.0	88	0.00
83 T	tert-Butylbenzene	2.692	2.922	-8.5	91	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.307	-5.8	90	0.00
85 T	sec-Butylbenzene	4.020	4.191	-4.3	89	0.00
86 T	p-Isopropyltoluene	3.392	3.649	-7.6	90	0.00
87 T	1,3-Dichlorobenzene	1.726	1.826	-5.8	92	0.00
88 T	1,4-Dichlorobenzene	1.716	1.826	-6.4	94	0.00
89 T	n-Butylbenzene	3.071	3.272	-6.5	90	0.00
90 T	Hexachloroethane	0.647	0.680	-5.1	92	0.00
91 T	1,2-Dichlorobenzene	1.505	1.650	-9.6	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.106	0.127	-19.8	101	0.00
93 T	1,2,4-Trichlorobenzene	0.885	1.098	-24.1	102	0.00
94 T	Hexachlorobutadiene	0.572	0.657	-14.9	99	0.00
95 T	Naphthalene	1.644	2.174	-32.2#	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010423\
Data File : VY012072.D
Acq On : 04 Jan 2023 16:42
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jan 05 00:42:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y122022S.M
Quant Title : SW846 8260
QLast Update : Tue Dec 20 12:37:37 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.749	0.959	-28.0	# 105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6