

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010725\
 Data File : VY020790.D
 Acq On : 07 Jan 2025 10:30
 Operator : SY/MD
 Sample : VY0107SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0107SBL01

Quant Time: Jan 08 00:52:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.732	168	186018	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.634	114	313904	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.432	117	251063	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.365	152	89480	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.085	65	81052	51.263	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.520%
35) Dibromofluoromethane	7.658	113	94174	47.887	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.780%
50) Toluene-d8	10.128	98	357888	52.204	ug/l	0.02
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.400%
62) 4-Bromofluorobenzene	12.420	95	100621	39.746	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	79.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010725\
Data File : VY020790.D
Acq On : 07 Jan 2025 10:30
Operator : SY/MD
Sample : VY0107SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0107SBL01

Quant Time: Jan 08 00:52:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

