

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010825\
 Data File : VY020803.D
 Acq On : 08 Jan 2025 11:50
 Operator : SY/MD
 Sample : Q1029-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-02-01072025

Quant Time: Jan 09 00:19:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 01:31:46 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.719	168	1517	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.621	114	2536	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.432	117	925	50.000	ug/l	# 0.01
72) 1,4-Dichlorobenzene-d4	13.364	152	335	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.073	65	977	75.772	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.540%
35) Dibromofluoromethane	7.652	113	955	60.109	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	120.220%
50) Toluene-d8	10.121	98	2405	43.423	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.840%
62) 4-Bromofluorobenzene	12.419	95	239	11.685	ug/l	0.01
Spiked Amount	50.000	Range	29 - 146	Recovery	=	23.380%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010825\
Data File : VY020803.D
Acq On : 08 Jan 2025 11:50
Operator : SY/MD
Sample : Q1029-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-02-01072025

Quant Time: Jan 09 00:19:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010625S.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 01:31:46 2025
Response via : Initial Calibration

