

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
 Data File : VY012110.D
 Acq On : 09 Jan 2023 15:00
 Operator : KP/MD
 Sample : 01064-02
 Misc : 6.20g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S4-02

Quant Time: Jan 10 01:47:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	162171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	276970	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	264718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	122825	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	104505	57.349	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.700%
35) Dibromofluoromethane	7.722	113	84943	46.660	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.320%
50) Toluene-d8	10.178	98	330607	46.010	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.020%
62) 4-Bromofluorobenzene	12.483	95	107918	42.967	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
Data File : VY012110.D
Acq On : 09 Jan 2023 15:00
Operator : KP/MD
Sample : 01064-02
Misc : 6.20g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S4-02

Quant Time: Jan 10 01:47:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

