

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
 Data File : VY012113.D
 Acq On : 09 Jan 2023 16:10
 Operator : KP/MD
 Sample : 01064-05
 Misc : 5.29g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-13-02

Quant Time: Jan 10 01:48:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	165896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	285268	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	270266	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	125926	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	107789	57.823	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.640%
35) Dibromofluoromethane	7.716	113	88189	47.034	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.060%
50) Toluene-d8	10.179	98	342760	46.314	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.620%
62) 4-Bromofluorobenzene	12.483	95	112423	43.459	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.920%
Target Compounds					Qvalue	
16) Acetone	3.954	43	20514	46.781	ug/l	97
25) 2-Butanone	6.984	43	4839	8.364	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
Data File : VY012113.D
Acq On : 09 Jan 2023 16:10
Operator : KP/MD
Sample : 01064-05
Misc : 5.29g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-13-02

Quant Time: Jan 10 01:48:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

