

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
 Data File : VY012122.D
 Acq On : 09 Jan 2023 19:41
 Operator : KP/MD
 Sample : 01084-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-2

Quant Time: Jan 10 01:50:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	185340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	313015	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	298353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	139679	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	108096	51.905	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.800%
35) Dibromofluoromethane	7.716	113	93891	45.636	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.280%
50) Toluene-d8	10.179	98	372955	45.927	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.860%
62) 4-Bromofluorobenzene	12.483	95	124873	43.993	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
Data File : VY012122.D
Acq On : 09 Jan 2023 19:41
Operator : KP/MD
Sample : 01084-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-2

Quant Time: Jan 10 01:50:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

