

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
 Data File : VY012127.D
 Acq On : 09 Jan 2023 21:38
 Operator : KP/MD
 Sample : 01057-06
 Misc : 7.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC-5

Quant Time: Jan 10 01:52:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	178929	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	283125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	229070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85726	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	79254	39.419	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.840%
35) Dibromofluoromethane	7.716	113	78802	42.345	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.700%
50) Toluene-d8	10.179	98	325462	44.310	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.620%
62) 4-Bromofluorobenzene	12.483	95	84751	33.010	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
Data File : VY012127.D
Acq On : 09 Jan 2023 21:38
Operator : KP/MD
Sample : 01057-06
Misc : 7.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC-5

Quant Time: Jan 10 01:52:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

