



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY010923\
 Data File : VY012102.D
 Acq On : 09 Jan 2023 11:10
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Title : SW846 8260
 Last Update : Sat Jan 07 01:46:48 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	11252	PASS
75	95	30	60	53.2	34043	PASS
95	95	100	100	100.0	64040	PASS
96	95	5	9	6.6	4252	PASS
173	174	0.00	2	1.3	760	PASS
174	95	50	100	90.2	57789	PASS
175	174	5	9	7.9	4548	PASS
176	174	95	101	96.5	55771	PASS
177	176	5	9	6.7	3709	PASS