

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011023\
 Data File : VY012135.D
 Acq On : 10 Jan 2023 12:34
 Operator : KP/MD
 Sample : 01076-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BU-03-010923

Quant Time: Jan 11 03:08:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	179466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	298900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	283001	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	127148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	102638	50.897	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.800%
35) Dibromofluoromethane	7.722	113	90427	46.028	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.060%
50) Toluene-d8	10.185	98	354269	45.686	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.380%
62) 4-Bromofluorobenzene	12.483	95	114726	42.326	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.660%
Target Compounds						
16) Acetone	3.954	43	7087	14.940	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011023\
Data File : VY012135.D
Acq On : 10 Jan 2023 12:34
Operator : KP/MD
Sample : 01076-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BU-03-010923

Quant Time: Jan 11 03:08:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

