

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011023\
 Data File : VY012142.D
 Acq On : 10 Jan 2023 15:20
 Operator : KP/MD
 Sample : 01090-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-3-010923

Quant Time: Jan 11 03:09:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
 Quant Title : SW846 8260
 QLast Update : Sat Jan 07 01:46:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	59719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	97410	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	92490	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	40723	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	35880	53.469	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.940%
35) Dibromofluoromethane	7.722	113	30805	48.113	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.220%
50) Toluene-d8	10.179	98	115261	45.610	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.220%
62) 4-Bromofluorobenzene	12.483	95	34604	39.174	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY011023\
Data File : VY012142.D
Acq On : 10 Jan 2023 15:20
Operator : KP/MD
Sample : 01090-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-3-010923

Quant Time: Jan 11 03:09:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y010623S.M
Quant Title : SW846 8260
QLast Update : Sat Jan 07 01:46:48 2023
Response via : Initial Calibration

